



**Corporate Air Parts**  
**FAA Repair Station # OOCR964K**  
7641 Densmore Ave, Van Nuys, CA 91406-2043

**Certificate of Calibration**

Print Date 4/2/2008

Certificate Number:07-00571

<p><b>Gage ID</b> 0020-CAPS  <b>Gage S/N</b> N0098  <b>Description</b> Pressure, Digital 0-10 psi  <b>Asset No.</b>  <b>Model No.</b> 475-4-FM  <b>Unit of Meas.</b> PSI  <b>Manufacturer</b> Dwyer  <b>Cal. Date</b> 6/5/2007  <b>Next Due</b> 6/5/2008  <b>Cal. Freq.</b> 1.00                      Years  <b>Location</b> Corporate Air Parts, Inc.</p>	<p align="center"><b>Approved</b> Yes  <b>As Found Condition</b> In  <b>Customer Notes:</b></p> <p align="center">Customer: <b>Corporate Air Parts, Inc.</b></p> <p align="center"><b>Customer P/O: SHOP</b></p>
--	--

**Certification Statement**

It is hereby certified that the above described instrument conforms to the original manufacturer's specifications and has been calibrated using standards whose accuracies are traceable to the NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY (NIST) within the limitations of the Institute Calibration Services or have been derived from accepted values of natural physical constants or have been derived by the ratio type of self calibration techniques.

**Findings**

**Environmental Conditions**

<p>In Calibration</p>	<p><b>Temperature</b> 68.96  <b>Humidity</b> 42.1  <b>Pressure</b> 29.90  <b>Customer P/O:</b> SHOP</p>
-----------------------	---

<b>Check Point</b>	1. (1.0 psi)	<b>Uncertainty</b>		<b>Units</b> psi		<b>Type</b> V
<b>Limited Use?</b>	No	<b>Minimum</b>	0.500	<b>Nominal</b>	1.000	<b>Maximum</b> 1.500
<b>Ref Type</b>	Pressure Calibrator	<b>Before</b>	0.990	<b>Accuracy</b>	-0.010	<b>Fail Before</b> No
		<b>After</b>	0.990	<b>Accuracy</b>	-0.010	<b>Fail After</b> No
<b>Gage ID of Standard</b>	0004-CAPS	<b>Gage S/N</b>	5843	<b>Model No.</b>	PC6	<b>NIST No.</b> 08006216034
<b>Std Due Date</b>	4/17/2008					

<b>Check Point</b>	2. (2.0 psi)	<b>Uncertainty</b>		<b>Units</b> psi		<b>Type</b> V
<b>Limited Use?</b>	No	<b>Minimum</b>	1.500	<b>Nominal</b>	2.000	<b>Maximum</b> 2.500
<b>Ref Type</b>	Pressure Calibrator	<b>Before</b>	2.080	<b>Accuracy</b>	0.080	<b>Fail Before</b> No
		<b>After</b>	2.080	<b>Accuracy</b>	0.080	<b>Fail After</b> No
<b>Gage ID of Standard</b>	0004-CAPS	<b>Gage S/N</b>	5843	<b>Model No.</b>	PC6	<b>NIST No.</b> 08006216034
<b>Std Due Date</b>	4/17/2008					

<b>Check Point</b>	3. (4.0 psi)	<b>Uncertainty</b>		<b>Units</b> psi		<b>Type</b> V
--------------------	--------------	--------------------	--	------------------	--	---------------

